

# Rat RT<sup>2</sup> RNA QC PCR Array

Cat. no. 330291 PARN-999ZA

## For real-time PCR applications

The Rat RT<sup>2</sup> RNA QC PCR Array, when used together with the RT<sup>2</sup> First Strand Kit (cat. no. 330401), is designed to assess the quality of rat RNA samples before characterization with RT<sup>2</sup> Profiler PCR Arrays or RT<sup>2</sup> qPCR Primer Assays in real-time PCR-based gene expression analysis. It contains a number of PCR controls, listed below, that test for RNA integrity, inhibitors of reverse transcription and PCR amplification, and genomic and general DNA contamination. These are all factors that could lead to unreliable results in SYBR<sup>®</sup> Green real-time PCR. Use of the RT<sup>2</sup> RNA QC PCR Array and the RT<sup>2</sup> First Strand Kit enables testing of the quality of RNA samples, thereby ensuring that RT<sup>2</sup> SYBR Green qPCR Mastermix and RT<sup>2</sup> Profiler PCR Arrays or RT<sup>2</sup> qPCR Primer Assays are not used with substandard samples. For full protocol details, consult the *RT<sup>2</sup> Profiler PCR Array Handbook*.

## Contents

Product	Contents	Cat. no.
RT <sup>2</sup> RNA QC PCR Array Format A	96-well RT <sup>2</sup> RNA QC PCR Array, 12 Optical Thin-Wall 8-Cap Strips; Suitable for use with the following real-time cyclers: Applied Biosystems <sup>®</sup> models 5700, 7000, 7300, 7500, 7700, 7900HT, ViiA <sup>™</sup> 7 (96-well block); Bio-Rad <sup>®</sup> models iCycler <sup>®</sup> , iQ <sup>™</sup> 5, MyiQ <sup>™</sup> , MyiQ2; Bio-Rad/MJ Research Chromo4 <sup>™</sup> ; Eppendorf <sup>®</sup> MasterCycler <sup>®</sup> ep realplex models 2, 2s, 4, 4s; Stratagene <sup>®</sup> models Mx3005P <sup>®</sup> , Mx3000P <sup>®</sup> ; Takara: TP-800	330291
RT <sup>2</sup> RNA QC PCR Array Format C	96-well RT <sup>2</sup> RNA QC PCR Array, Optical Adhesive Film; Suitable for use with the following real-time cyclers: Applied Biosystems models 7500 (Fast block), 7900HT (Fast block), StepOnePlus <sup>™</sup> , ViiA 7 (Fast block)	330291
RT <sup>2</sup> RNA QC PCR Array Format D	96-well RT <sup>2</sup> RNA QC PCR Array, 12 Optical Thin-Wall 8-Cap Strips; Suitable for use with the following real-time cyclers: Bio-Rad CFX96 <sup>™</sup> ; Bio-Rad/MJ Research models DNA Engine Opticon <sup>®</sup> , DNA Engine Opticon 2; Stratagene Mx4000 <sup>®</sup>	330291



Product	Contents	Cat. no.
RT <sup>2</sup> RNA QC PCR Array Format E	384-well RT <sup>2</sup> RNA QC PCR Array, Optical Adhesive Film; Suitable for use with the following real-time cyclers: Applied Biosystems models 7900HT (384-well block), ViiA 7 (384-well block); Bio-Rad CFX384™	330291
RT <sup>2</sup> RNA QC PCR Array Format F	96-well RT <sup>2</sup> RNA QC PCR Array, Optical Adhesive Film; Suitable for use with the following real-time cyclers: Roche® LightCycler® 480 (96-well block)	330291
RT <sup>2</sup> RNA QC PCR Array Format G	384-well RT <sup>2</sup> RNA QC PCR Array, Optical Adhesive Film; Suitable for use with the following real-time cyclers: Roche: LightCycler 480 (384-well block)	330291
RT <sup>2</sup> RNA QC PCR Array Format R	Rotor-Disc® 100 RT <sup>2</sup> RNA QC PCR Array, Suitable for use with the following real-time cyclers: QIAGEN® Rotor-Gene® Q; Rotor-Gene 6000; other Rotor-Gene cyclers	330291

## Shipping and storage

RT<sup>2</sup> RNA QC PCR Array Formats A, C, D, E, F, G, and R are shipped at room temperature (15–25°C) or on ice, depending on the destination and accompanying products. All RT<sup>2</sup> RNA QC PCR Array Formats should be stored at –20°C upon arrival. When stored properly at –20°C, RT<sup>2</sup> RNA QC PCR Arrays are stable for up to 6 months after delivery.

**Note:** Open the package and store the products appropriately immediately on receipt.

## 96-well RT<sup>2</sup> RNA QC PCR Array (Formats A, C, D, F) layout

Row	Content	Material to add	Control type
A	ACTB	cDNA, mastermix	Housekeeping gene
B	HPRT1	cDNA, mastermix	Housekeeping gene
C	RTC	cDNA, mastermix	Reverse transcription control
D	PPC	cDNA, mastermix	Positive PCR control
E	GDC	cDNA, mastermix	Rat genomic DNA contamination control
F	NRT	RNA, mastermix	No reverse transcription control
G	PPC	Water, mastermix	Positive PCR control
H	NTC	Water, mastermix	No template control

### 384-well RT<sup>2</sup> RNA QC PCR Array (Formats E, G) layout

Row	Content	Material to add	Control type
A, B	ACTB	cDNA, mastermix	Housekeeping gene
C, D	HPRT1	cDNA, mastermix	Housekeeping gene
E, F	RTC	cDNA, mastermix	Reverse transcription control
G, H	PPC	cDNA, mastermix	Positive PCR control
I, J	GDC	cDNA, mastermix	Rat genomic DNA contamination control
K, L	NRT	RNA, mastermix	No reverse transcription control
M, N	PPC	Water, mastermix	Positive PCR control
O, P	NTC	Water, mastermix	No template control

### Rotor-Disc 100 RT<sup>2</sup> RNA QC PCR Array (Format R) layout

Well	Content	Material to add	Control type
1–12	ACTB	cDNA, mastermix	Housekeeping gene
13–24	HPRT1	cDNA, mastermix	Housekeeping gene
25–36	RTC	cDNA, mastermix	Reverse transcription control
37–48	PPC	cDNA, mastermix	Positive PCR control
49–60	GDC	cDNA, mastermix	Rat genomic DNA contamination control
61–72	NRT	RNA, mastermix	No reverse transcription control
73–84	PPC	Water, mastermix	Positive PCR control
85–96	NTC	Water, mastermix	No template control

## Quality control parameters for 96-well and 384-well formats

Control type	Pass test result*	Fail test result*	Pass test result*	Fail test result*
	Formats A, C, D, F	Formats A, C, D, F	Formats E, G	Formats E, G
Reverse transcription control	$C_T^C - C_T^G$ value < 5	$C_T^C - C_T^G$ value > 5	$C_T^{E,F} - C_T^{M,N}$ value < 5	$C_T^{E,F} - C_T^{M,N}$ value > 5
Positive PCR control	$C_T^G = 20 \pm 2$ $C_T^D - C_T^G < 3$	$C_T^G > 22$ $C_T^D - C_T^G > 3$	$C_T^{M,N} = 20 \pm 2$ $C_T^{G,H} - C_T^{M,N} < 3$	$C_T^{M,N} > 22$ $C_T^{G,H} - C_T^{M,N} > 3$
Genomic DNA contamination control	$C_T^E > 35$	$C_T^E < 35$	$C_T^{I,J} > 35$	$C_T^{I,J} < 35$
No reverse transcription control	$C_T^F$ value $\geq 35$ or N/A	$C_T^F$ value < 35	$C_T^{K,L}$ value $\geq 35$ or N/A	$C_T^{K,L}$ value < 35
No template control	$C_T^H$ value $\geq 35$ or N/A	$C_T^H$ value < 35	$C_T^{O,P}$ value $\geq 35$ or N/A	$C_T^{O,P}$ value < 35

\* Superscript indicates row designation.

## Quality control parameters for Rotor-Disc format

Control type	Pass test result*	Fail test result*
	Formats R	Format R
Reverse transcription control	$C_T^{25 \text{ to } 36} - C_T^{73 \text{ to } 84}$ value < 5	$C_T^{25 \text{ to } 36} - C_T^{73 \text{ to } 84}$ value > 5
Positive PCR control	$C_T^{73 \text{ to } 84} = 14 \pm 2$ $C_T^{37 \text{ to } 48} - C_T^{73 \text{ to } 84} < 3$	$C_T^{73 \text{ to } 84} > 16$ $C_T^{37 \text{ to } 48} - C_T^{73 \text{ to } 84} > 3$
Genomic DNA contamination control	$C_T^{49 \text{ to } 60} > 33$	$C_T^{49 \text{ to } 60} < 33$
No reverse transcription control	$C_T^{61 \text{ to } 72}$ value $\geq 33$ or N/A	$C_T^{61 \text{ to } 72}$ value < 33
No template control	$C_T^{85 \text{ to } 96}$ value $\geq 33$ or N/A	$C_T^{85 \text{ to } 96}$ value < 33

\* Superscript indicates well designation.

## Related products

Product	Contents	Cat. no.
RT <sup>2</sup> Profiler PCR Array	Pathway- or disease-specific primer assays in 96-well, 384-well, Rotor-Disc 100, or 96 x 96 format	Varies
RT <sup>2</sup> First Strand Kit (12)	Enzymes and reagents for cDNA synthesis	330401
RT <sup>2</sup> SYBR Green qPCR Mastermix (2)*	For 2 x 96 assays in 96-well plates; suitable for use with real-time cyclers that do not require a reference dye, including: Bio-Rad models CFX96, CFX384, DNA Engine Opticon 2; Bio-Rad/MJ Research Chromo4; Roche LightCycler 480 (96-well and 384-well); all other cyclers	330500
RT <sup>2</sup> SYBR Green ROX qPCR Mastermix (2)*	For 2 x 96 assays in 96-well plates; suitable for use with the following real-time cyclers: Applied Biosystems models 5700, 7000, 7300, 7500 [Standard and FAST], 7700, 7900HT 96-well block [Standard and FAST] and 384-well block, StepOnePlus; Eppendorf Mastercycler ep realplex models 2, 2S, 4, 4S; Stratagene models Mx3000p, Mx3005p, Mx4000; Takara TP-800	330520
RT <sup>2</sup> SYBR Green Fluor qPCR Mastermix (2)*	For 2 x 96 assays in 96-well plates; suitable for use with the following real-time cyclers: Bio-Rad models iCycler, iQ5, MyiQ, MyiQ2	330510
RT <sup>2</sup> SYBR Green ROX FAST Mastermix (2)*	For 2 x 96 assays in 96-well plates/100-well Rotor-Discs; suitable for use with the Rotor-Gene Q and other Rotor-Gene cyclers	330620

\* Larger kit sizes available; please inquire.

The RT<sup>2</sup> RNA QC PCR Array is intended for molecular biology applications. This product is not intended for the diagnosis, prevention, or treatment of a disease.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at [www.qiagen.com](http://www.qiagen.com) or can be requested from QIAGEN Technical Services or your local distributor.

Trademarks: QIAGEN<sup>®</sup>, Rotor-Gene<sup>®</sup>, Rotor-Disc<sup>®</sup> (QIAGEN Group); Roche<sup>®</sup>, LightCycler<sup>®</sup> (Roche Group); Applied Biosystems<sup>®</sup>, ROX<sup>™</sup>, StepOnePlus<sup>™</sup>, SYBR<sup>®</sup>, ViiA<sup>™</sup> (Life Technologies Corporation); Eppendorf<sup>®</sup>, Mastercycler<sup>®</sup> (Eppendorf AG); Stratagene<sup>®</sup>, Mx3005P<sup>®</sup>, Mx3000P<sup>®</sup>, Mx4000<sup>®</sup> (Agilent Technologies); Bio-Rad<sup>®</sup>, iCycler<sup>®</sup>, Chromo4<sup>™</sup>, CFX96<sup>™</sup>, DNA Engine Opticon<sup>®</sup>, CFX384<sup>™</sup>, iQ<sup>™</sup>, MyiQ<sup>™</sup> (Bio-Rad Laboratories, Inc.).

1072634 05/2012 © 2012 QIAGEN, all rights reserved.

[www.qiagen.com](http://www.qiagen.com)

**Canada** ■ 800-572-9613

**China** ■ 8621-3865-3865

**Denmark** ■ 80-885945

**Australia** ■ 1-800-243-800

**Finland** ■ 0800-914416

**Austria** ■ 0800/281010

**France** ■ 01-60-920-930

**Belgium** ■ 0800-79612

**Germany** ■ 02103-29-12000

**Brazil** ■ 0800-557779

**Hong Kong** ■ 800 933 965

**Ireland** ■ 1800 555 049

**Italy** ■ 800-787980

**Japan** ■ 03-6890-7300

**Korea (South)** ■ 080-000-7145

**Luxembourg** ■ 8002 2076

**Mexico** ■ 01-800-7742-436

**The Netherlands** ■ 0800 0229592

**Norway** ■ 800-18859

**Singapore** ■ 1800-742-4368

**Spain** ■ 91-630-7050

**Sweden** ■ 020-790282

**Switzerland** ■ 055-254-22-11

**UK** ■ 01293-422-911

**USA** ■ 800-426-8157



---

**Sample & Assay Technologies**